

«УТВЕРЖДАЮ»

Генеральный директор

ООО «Медтроник»

Пясунова Е.В.



Фотографические изображения

«Электрокардиостимулятор имплантируемый Vitatron A60 DR модель A60A1»

Производства:

«Витатрон Холдинг Б.В.», Нидерланды


Vitatron Holding BV., Endepolsdomein 5, 6229 GW Maastricht, The Netherlands



A60 DR

vitatron

Reduced VP™ +




SN 350001196G

2017-11-28

2016-06

CE0344



DDDR

150 ms
120 ms
A-V interval (paced / sensed)

0.5 mV
Atrial sensitivity

3.5 V
0.4 ms
Atrial amplitude / pulse width

AUTO
Sensing polarity

85 min
Magnet rate

130 min
60 min
Upper tracking rate / lower rate

230 ms
Ventricular refractory period

2.8 mV
Ventricular sensitivity

3.5 V
0.4 ms
RV amplitude / pulse width

AUTO
Pacing polarity

AUTO
PVARP

Vitatron Holding B.V.
Endepolderdijk 5
6229 QW Maastricht
The Netherlands




Manufactured in
Singapore
Product of Singapore

STERILIZED BY GAMMA RAY

A60 DR

vitatron

Reduced VP™ +





SN 350001196G

2017-11-28

2016-06

CE0344




SN 350001196G

2017-11-28

2016-06

CE0344



Manufactured in
Singapore
Product of Singapore



vitatron

A-series

A60 DR A60A1

Pacemaker Manual • Schrittmacherhandbuch •
Εγχειρίδιο βηματοδότη • Умрулати а pacemaker
naszalatához • Instrukcja obsługi stymulatora serca
• Páručka ke kardiostimulátoru • Руководство по
3HC • Páručka ku kardiostimulátoru

vitatron

Terms of Limited Warranty

- This Limited Warranty provides the following assurance to a patient who receives the Vitatron pulse generator:
 - Should the pulse generator fail to function within normal parameters at any time due to a defect in materials or workmanship, Vitatron will repair or replace the pulse generator at no cost to the patient. This warranty does not cover the pulse generator if it is not considered a defect in materials or workmanship.
 - As credit as defined in Subsection A.2.2, will be issued if the pulse generator is replaced within 60 months from the date of purchase, in accordance with the pulse generator's Recommended Replacement Time Indicator. Each warranty period commences on the implant date. The warranty on battery depletion is made void if the pulse generator has been programmed to output values higher than the standard factory settings for an uninterrupted period of longer than 3 months.
 - The credit is equal to the purchase price and is issued against the purchase of any Vitatron pulse generator requested for a replacement, or, at the discretion of Vitatron, if a functionally comparable replacement Vitatron pulse generator at no charge. Normally the credit will be issued only to the hospital or physician purchasing the replacement pulse generator.
- To qualify for the Limited Warranty, the following conditions must be met:
 - The pulse generator must be implanted prior to its "Use By" date.
 - The pulse generator for which credit is sought must be returned to Vitatron in a stockproof package within 30 days of its removal and shall then become the property of Vitatron. Furthermore, a "Request for analysis" form should be completed and returned with the pulse generator.
- This Limited Warranty is limited to the express terms, in particular:
 - Vitatron shall not be liable to any person for any medical expenses or any direct or consequential damages resulting from implant or removal of any pulse generator or caused by any defect, failure or malfunction of any pulse generator, whether a claim for such damages is based upon warranty, contract, tort or otherwise. The exclusive remedy for any defect, failure or malfunction of any pulse generator shall be as stated in Subsection A.2.1.
 - This Limited Warranty is made only to the patient in whom the pulse generator was originally implanted and is expressly in lieu of any implied warranty of merchantability or fitness for a particular purpose, whether arising from statute, common law, custom or otherwise. The remedies set forth in this Limited Warranty shall be the exclusive remedies available to any person. No person has any authority to bind Vitatron to any representation, condition or warranty except as contained in this Limited Warranty.
 - This Limited Warranty is only valid as to the pulse generator and is not applicable, in whole or part, to any other Vitatron product, including, but not limited to, Vitatron pacing leads.

vitatron

.....

A60 DR

Reduced VP™ +



A60A1

M954387A001 A - M954379A001 A

SN 350001196G



2017-11-28

YYYY-MM-DD

REF A60A1

DR

IS-1 BI

IS-1 UNI

AUTO



(01)00643169232037(17)171128(21)350001196G



vitatron

A60 DR

Reduced VP™ +



A60A1

M954387A001 A - M954379A001 A

SN

350001196G



2017-11-28

YYYY-MM-DD

DR

IS-1 BI

IS-1 UNI

AUTO

en Pacemaker / de Herzschrittmacher / el Βηματοδότης /
hu Pacemaker / pl Stymulator / cs Kardiostimulátor /
ru Электрокардиостимулятор / sk Kardiostimulátor

PIN: A60A100002

Генеральный директор
ООО «Медтроник»
Плясунова Е.В.

5 листов

