


УТВЕРЖДАЮ
Специалист по регистрации
ООО «Джонсон и Джонсон»
Прохорова Л.В.
« 22 » ноября 2016 г.



Фотографические изображения медицинского изделия: «Система VAPR для артроскопических вмешательств»

производства: DePuy Mitek, Inc., 325 Paramount Drive, Raynham MA, 02767, USA

Содержание:

Система VAPR для артроскопических вмешательств: Рукоятка (Артикул: 225002)	2
Система VAPR для артроскопических вмешательств: Педаль (Артикул: 225023)	5
Система VAPR для артроскопических вмешательств: Педаль (Артикул: 227214)	10
Система VAPR для артроскопических вмешательств: Генератор (VAPR VUE, Артикул: 225024)	12
Система VAPR для артроскопических вмешательств: Генератор (VAPR 3, Артикул: 225021)	15



DePuy Mitek
 VAPR® Handpiece
 REUSABLE
 REF 225002
 LOT 1303183
 QTY 1
 119125-G8
 11/17
 DePuy Mitek, Inc. 325 Boulevard Drive, Redwood, VA 07871 USA, Tel: 800.345-4035
 DePuy Mitek Ltd, 9, Znamenskaya Street, London, E1 7AN, UK
 Manufactured for DePuy Mitek Ltd, 9, Znamenskaya Street, London, E1 7AN, UK
 CE
 119125
 MADE IN UK
 119125-G8
 11/17



VAPR® Handpiece
DePuy Mitek
 REF 225002

Demonstration ONLY.
 Not for human use!
 Демонстрационная операция
 Не для использования
 на человеке



DePuy Mitek
 VAPR® Handpiece
 CE 0086
 Manufactured for DePuy Mitek by
 Gyrex Medical P/N: 119125-G8
 Rev: G Revised 01/17
 P/N: 10096
 © 2004-2012 DePuy Mitek





Система VAPR для артроскопических вмешательств: Педаль (Артикул: 225023)













REF 792025

LOT: 1902057

MANUFACTURED: 01/2011

DePuy Mitek,
325 Paramount Drive,
Raynham,
MA, 02767, USA
Tel (800)356-4835

EC REP

DePuy International Ltd,
St Anthony's Rd, Leeds,
LS11 8DT, ENGLAND



IP68

MADE IN USA



192164-AC

VAPR® VUE Wireless Footswitch p/no. 227214

Manufactured for
DePuy Mitek by:
Gyrus Medical Ltd,
Fortran Road,
St Mellons, Cardiff,
CF3 0LT, UK

CE
0086



L03903246TXA

FCC ID: VPA-SP-831

THIS DEVICE COMPLIES WITH PART 15 OF
THE FCC RULES. OPERATION IS SUBJECT TO
THE FOLLOWING TWO CONDITIONS:
1) THIS DEVICE MAY NOT CAUSE HARMFUL INTERFERENCE AND
2) THIS DEVICE MUST ACCEPT ANY INTERFERENCE RECEIVED,
INCLUDING INTERFERENCE THAT MAY CAUSE UNDESIRABLE
OPERATION.





RECEIVED IN CHINA
MANUFACTURED IN CHINA

153037-24H

SERVICE USE ONLY

Grounding reliability can only be achieved when this equipment is connected to an equivalent receptacle marked "Hospital only" or "Hospital grade"

DePuy Mitek,
325 Paramount Drive,
Raynham,
MA, 02767, USA
Tel (800)356-4835

EC REP

DePuy International Ltd,
St Anthony's Rd, Leeds,
LS11 8DT, England

WARNING: RISK OF FIRE
REPLACE FUSE AS MARKED

SN 1322045 2013-10

CLASSIFIED
UL US
21FA

REF: 225024

MADE IN UK

100-120/230V~
50/60Hz 1000VA

100-120/230V~
50/60Hz 1000VA

Manufactured for DePuy
Mitek by:
Gyrus Medical Ltd,
Fortran Road, St Mellons,
Cardiff, CF3 0LT, UK

CE
0086

DePuy Mitek,
 325 Paramount Drive,
 Raynham,
 MA, 02767, USA
 Tel (800)356-4835

EC REP
 DePuy International Ltd,
 St Anthony's Rd, Leeds,
 LS11 8DT, England

WARNING: RISK OF FIRE
REPLACE FUSE AS MARKED

100-120/230V~
 50/60Hz 1000VA



SN 1322045 2013-10

CLASSIFIED
UL
US
 21FA

Manufactured for DePuy
 Mitek by:
 Gyrus Medical Ltd,
 Fortran Road, St Mellons,
 Cardiff, CF3 0LT. UK

CLASS 1	V1/2/3 OUTPUT	COAG OUTPUT	CP OUTPUT
IEC 60601-1	300-450KHz	300-450KHz	360KHz
INT. 10s ON/ 30s OFF	260W/330Ω	200W/28Ω	400W/68Ω



REF: 225024

MADE IN UK

CE
 0086

FOR MEDICAL USE ONLY


...g reliability can only be
 ...when this equipment is
 ...ected to an equivalent
 ...le marked "Hospital only"
 ...or "Hospital grade"





137002-F

REF: 225021	SERIAL No:
MANUFACTURED:	
DePuy Mitek, 325 Paramount Drive, Raynham, MA, 02767, USA Tel (800)356-4835	EU Representative: DePuy International Ltd, St Anthony's Rd, Beeston, Leeds, LS11 8DT, UK Tel +44 113270 0461



CLASS 1
IEC 60601-1
340-450 KHz
260W/160Ω
INT. 10s/30s

**WARNING: RISK OF FIRE
REPLACE FUSE AS MARKED**

100-120/220-240V 50/60Hz
500VA 2 x T6.3A

Manufactured for
DePuy Mitek by:
Gyrus Medical Ltd,
Fortran Road,
St Mellons, Cardiff,
CF3 0LT, UK
Tel +44 29 2077 6300

Сшито и заверено печатью

16 листов

Прохорова Л.В.

