

digene HC2 HPV DNA Test — Quick-Start Guide

The *digene* HPV Test



1. **Preparation of the test kit:** The test kit should be stored at 2-8°C. Before use, the kit should be allowed to reach room temperature. The test kit should be opened in a clean, dry, and well-ventilated area. The test kit should be handled with care to avoid contamination.

2. **Sample collection:** The sample should be collected using a sterile swab. The swab should be inserted into the sample collection device and rotated to collect the sample. The sample should be collected from the cervix or the vaginal canal.

3. **Sample storage:** The sample should be stored in the sample collection device at 2-8°C. The sample should be used within 48 hours of collection.

4. **Test procedure:** The test procedure involves the following steps:

1. Preparation of the test kit.
2. Sample collection.
3. Sample storage.
4. Test procedure.
5. Result interpretation.

5. **Result interpretation:** The result of the test is interpreted based on the presence or absence of HPV DNA. A positive result indicates the presence of HPV DNA, while a negative result indicates the absence of HPV DNA.

