

ФОТОГРАФИЧЕСКОЕ ИЗОБРАЖЕНИЕ

**Устройство для выбивания образцов
крови «Panthera-Puncher 9»
(изготовитель Wallac Oy, Финляндия).**

Оборудование поставляется в комплекте с принадлежностями:

1. Сенсорный экран с подставкой -1 штука.
2. Сборные головки перфоратора - до 10 штук.
3. Ножная педаль -1 штука.
4. Адаптеры для плашек -до 10 штук.
5. Внешний компьютер - 1 штука.
 - 5.1 Кабели для подключения – до 15 штук
 - 5.2. Установочный диск с документами для ПК
6. Манипулятор типа «Мышь» - 1 штука.
7. Внешняя клавиатура - 1 штука.
8. Комплект кабелей питания для подключения блока перфоратора.
 - 8.1 Кабель USB - 1 шт.
 - 8.2. Кабель CAN – 1 шт.
 - 8.3. Кабель 230-240 V - до 2 шт.
 - 8.4 Кабель 115 V – до 3 штук.
9. Установочное программное обеспечение на бумажных и дисковых носителях.
10. Руководство по эксплуатации.
11. Сертификат контроля качества, СЕ-марка.
12. Шестигранный ключ – 1 штука.

Вспомогательная система: насосная станция / Пункт Аварийной помощи






Секционный анализ - с регистрацией

Don't forget to sign.



Сопло для сверла



10862767

Pedal

Wipac Oy
Mustamäki 6, Turku, Finland



PerkinElmer

ga

klonirovany

61009227

Sample plate adapter_2081



Wallac Oy
Mustionkatu 6, Turku, Finland




PerkinElmer®


Адаптер для планшетов

Контроль температуры



Кабели для компьютеров
и периферии





Please refer to User Guide for detailed operating, maintenance and servicing instructions.

Установочный диск с документацией для ПК

Характеристики товара: цвет: черный



Вещи не брали





11241196

USB cable_ ferrite type A-B 2 m_
USB-218FS

Wallac Oy
Mustionkatu 6 Turku Finland



PerkinElmer

USB 2.0 2m

USB connection cable with
Ferrite cores



USB Type B M



USB Type A M

USB-218FS



7340004655770

USB

Кабель USB



Кабель CAN



11460011


Mains cable 240 V

Wälac Oy
Mustionkatu 6, Turku, Finland



PerkinElmer

P Group M



IEC C13

DEL-109M

CEE 177



734000400039

POWER

- Кабель 230-240 В



11460012
Mains cable 115 V

Wallac Oy
Mustionkatu 6, Turku, Finland

 PerkinElmer

POWER

Max 15A / 125V


IEC C13

DEL-100US

NEMA 5-15

7110001257002

Кабель 115V.



INSTRUMENT MANUAL

2081-0010-02
July 2012

2081-0010 Panthera- Puncher™ 9

Dried blood spot punching device



FOR IN VITRO DIAGNOSTIC USE




PerkinElmer

Римовгребто по мемиягтоуни

QUALITY CONTROL CERTIFICATE

PRODUCT INFORMATION


Name of product: Panthera-Puncher 9
Product type number: 2081-0010
Serial number: 20810048

CERTIFICATE

This is to certify that this instrument and its components and subunits have been tested according to the PerkinElmer, Wallac Quality System practices in compliance with ISO 9001 and ISO 13485 and that it fulfills the specified quality and performance values. More detailed information can be obtained from your PerkinElmer representative.

Date: 27.02.2013

Signature and
clarification:

 JUKKA NIITYVAARA

On behalf of the Quality Control Department

Wallac Oy Business Identity
Code: 0937168-4
Registered office: Turku



Сертификат качества

**DECLARATION OF CONFORMITY FOR CE-MARKING****Manufacturer's name and address**

Wallac Oy
Mustionkatu 6
FI-20750 Turku
Finland

Name of the device(s)

2081-0010 Panthera Puncher 9

We, Wallac Oy, hereby declare that the device(s) mentioned above comply with
2006/95/EC Low Voltage Directive

2004/108/EC Electromagnetic Compatibility Directive

98/79/EC In Vitro Diagnostic Medical device directive, as transposed to the Finnish legislation

The product(s) bears the CE mark indicating conformity with the provisions of these directives.

The product(s) are classified as follows according to the in vitro diagnostic medical devices
98/79/EC:

Other device (all devices except Annex II and self testing devices)

Conformity Assessment Procedure:

Self-declaration, Annex III, 98/79/EC

Common Technical Specifications used for demonstration of compliance:

EN61010-1 Safety requirements for electrical equipment for measurement, control and laboratory use

EN 61326-1 EMC requirements for electrical equipment for measurement, control and laboratory use

EN61010-2-101 Particular requirements for in vitro diagnostic (IVD) medical equipment

Date and place of issue

10. February 2012, Turku

Name, position and signature of authorized person

Tapio Kivimaa, Quality Manager

CE - marking

21

Внешний вид

11510014

Hexagon key 3mm



Wallac Oy
Mustonkatu 6, Turku, Finland



 PerkinElmer

Всего прошито и прошито
22 (Двадцать два) листа
Технический директор
ЗАО: Приборы
Грица Вадим Иванович

