

УТВЕРЖДАЮ

Генеральный директор
ООО «Компания «БиВи»

Носова Т.А.

«21» 07 2021 г.



ПО ДОВЕРЕННОСТИ
77 АГ 2759786

КУСАЙЛО Р.В.

Фотографические изображения изделия:

**Кассета с реагентами для количественного определения
нейронспецифической енолазы (Neuron-specific enolase (CLIA) / (NSE))
иммунохемилюминесцентным методом в клиническом образце на
анализаторах серии CL для диагностики in vitro**

Shenzhen Mindray Bio-Medical Electronics Co., Ltd. («Шэньчжэнь Майндрэй Био-
Медикал Электроникс Ко., Лтд.»), Китай

mindray
NSE

Neuron-specific enolase (CLIA)

Order Information

Catalog No.	Package Size
NSE111	2x50 tests
NSE112	2x100 tests
NSE113	2x30 tests

Intended Use

The CL-series NSE assay is a Chemiluminescent Immunoassay (CLIA) for the quantitative determination of neuron-specific enolase (NSE) in human serum.

Summary

The glycolytic enzyme enolase has several dimeric isoforms including and The α -subunit of enolase exists in various types of mammal tissues, whereas the β -subunit mainly exists in the heart and striated musculature. The β -subunit was found mainly in neurons, neuro-endocrine cells and in some tumors, so α -enolase is also known as neuron-specific enolase (NSE)¹. NSE was chosen as the first marker for monitoring the therapy and course of the disease in small cell bronchial carcinoma, of which 60-81% of cases show elevated NSE concentration^{2,3}. For responders to chemotherapy, elevated NSE level falls rapidly by the end of the first therapy cycle. By contrast, non-responders to therapy show constantly elevated level higher than the reference range^{4,5}. During remission, NSE value remains normal. In cases of relapse, NSE value will rise. For non-small cell

mindray
NSE

Two-dimensional Barcode of
Master Calibration Curve
LOT: 2020030100
EXP: 2022-02-03



NSE

IVD

CE



Neuron-specific enolase (CLIA)

2x100 tests

Contents: Ra: 2x6.6 mL, Rb: 2x6.3 mL

mindray



SHENZHEN MINDRAY BIO-MEDICAL ELECTRONICS CO., LTD.
Mindray Building, Keji 12th Road South, High-Tech Industrial Park,
Nanshan, Shenzhen, 518057, P.R. China
Shanghai International Holding Corp. GmbH(Europe)
Elffestraße 80, 20537 Hamburg, Germany

Neuron-specific enolase
(CLIA)

100 tests

Ra : 6.6 mL . Rb : 6.3 mL

NSE

IVD

LOT

2020030100



2022-02-03



mindray

Общий вид медицинского изделия (Кассета с реагентами для количественного определения нейронспецифической енолазы (Neuron-specific enolase (CLIA) / (NSE))
ИММУНОХЕМИЛЮМИНЕСЦЕНТНЫМ МЕТОДОМ в клиническом образце на анализаторах серии CL для диагностики in vitro, в варианте исполнения 2 x 100 тестов)

Всего прошито, пронумеровано и скреплено
печатью и подписью: 2 листа (-ов)

Генеральный директор _____ / Носова Т.А
М.П.

ПО ДОВЕРЕННОСТИ
77 АГ 2759786
КУСАЙЛО Р.В.

